

OSVEDČENIE O SÚPRAVE

ELISA-anti-MILK-II-G

Katalógové číslo	REF	E-245
Šarža	LOT	02 02 22
Dátum výroby		21.02.2022
Dátum expirácie		21.11.2022

Parametre výstupnej kontroly

Štandardy	protilátky anti IgG	
	Absorbancia pri 450 nm a 630 nm	
0 U.ml ⁻¹		0,005
6,25 U.ml ⁻¹		0,328
12,5 U.ml ⁻¹		0,556
25 U.ml ⁻¹		0,939
50 U.ml ⁻¹		1,288
100 U.ml ⁻¹		1,777

Protilátky triedy IgG

Kontrolné vzorky	Deklarovaná hodnota U.ml ⁻¹	Absorbancia pri 450 a 630 nm	Nameraná hodnota U.ml ⁻¹
negatívna 	< 9	0,136	1,90
pozitívna 	30 ± 6	1,066	31,79

ZÁVER:

Na základe výstupnej kontroly šarža výrobku vyhovuje špecifikácii - technologický reglement č.7 , Smernici 98/79/ES a NV SR č. 569/2001 Z. z. v platnom znení.

Dátum kontroly : 21.02.2022

Vedúca výrobného oddelenia:



Preveril:



BIOGEMA
výrobné družstvo, Košice
Garbiarska 2
040 01 KOŠICE ☎
tel.: 055/633 6753, 622 8131

O200809YB

