

OSVEDČENIE O SÚPRAVE

ELISA-anti-CANDIDA-II-M

Katalógové číslo	REF	E-353
Šarža	LOT	04 04 22
Dátum výroby		25.04.2022
Dátum expirácie		25.04.2023

Parametre výstupnej kontroly

Štandardy	protilátky anti IgM	
	Absorbancia pri 450 nm a 630 nm	
0 U.ml ⁻¹		0,010
6,25 U.ml ⁻¹		0,222
12,5 U.ml ⁻¹		0,388
25 U.ml ⁻¹		0,646
50 U.ml ⁻¹		1,038
100 U.ml ⁻¹		1,564

Protilátky triedy IgM

Kontrolné vzorky	Deklarovaná hodnota U.ml ⁻¹	Absorbancia pri 450 a 630 nm	Nameraná hodnota U.ml ⁻¹
negatívna 	< 12	0,070	1,542
pozitívna 	44 ± 8	0,922	41,755

ZÁVER:

Na základe výstupnej kontroly šarža výrobku vyhovuje špecifikácii - technologický reglement č.26 , Smernici 98/79/ES a NV SR č. 569/2001 Z. z. v platnom znení.

Dátum kontroly : 25.04.2022

Vedúca výrobného oddelenia:


BIOGEMA
výrobné družstvo, Košice
Garbiarska 2
040 01 KOŠICE
tel.: 055/633 6753, 622 8131

Preveril:



R210225RD

