









OSVEDČENIE O SÚPRAVE



ELISA-anti-MILK-II-A

Katalógové číslo	REF	E-244
Šarža	LOT	05 05 22
Dátum výroby		24.05.2022
Dátum expirácie		24.02.2023

Parametre výstupnej kontroly

Štandardy	protilátky anti IgA	
	Absorbancia pri 450 nm a 630 nm	
0 U.ml ⁻¹ 	0,010	
6,25 U.ml ⁻¹ 	0,221	
12,5 U.ml ⁻¹ 	0,407	
25 U.ml ⁻¹ 	0,648	
50 U.ml ⁻¹ 	1,076	
100 U.ml ⁻¹ 	1,618	

Protilátky triedy IgA

Kontrolné vzorky	Deklarovaná hodnota U.ml ⁻¹	Absorbancia pri 450 a 630 nm	Nameraná hodnota U.ml ⁻¹
negatívna 	< 9	0,043	1,11
pozitívna 	34 ± 7	0,760	31,14

ZÁVER:

Na základe výstupnej kontroly šarža výrobku vyhovuje špecifikácii - technologický reglement č.5 , Smernici 98/79/ES a NV SR č. 569/2001 Z. z. v platnom znení.

Dátum kontroly : 24.05.2022

Vedúca výrobného oddelenia:

Mihalikova

Preveril:

[Signature]

BIOGEMA
výrobné družstvo, Košice
Garbiarska 2
040 01 KOŠICE ☐
tel.: 055/633 6753, 622 8131

O210605UE

