

OSVEDČENIE O SÚPRAVE

ELISA-anti-MILK-II-A-200

Katalógové číslo	REF	E-244/200
Šarža	LOT	05 05 22
Dátum výroby		24.05.2022
Dátum expirácie		24.02.2023

Parametre výstupnej kontroly

Štandardy	protilátky anti IgA	
	Absorbancia pri 450 nm a 630 nm	
0 U.ml ⁻¹		0,007
6,25 U.ml ⁻¹		0,126
12,5 U.ml ⁻¹		0,225
25 U.ml ⁻¹		0,436
50 U.ml ⁻¹		0,725
200 U.ml ⁻¹		1,780

Protilátky triedy IgA

Kontrolné vzorky	Deklarovaná hodnota U.ml ⁻¹	Absorbancia pri 450 a 630 nm	Nameraná hodnota U.ml ⁻¹
negatívna 	< 9	0,039	1,42
pozitívna 	34 ± 7	0,519	30,94

ZÁVER:

Na základe výstupnej kontroly šarža výrobku vyhovuje špecifikácii - technologický reglement č.5 , Smernici 98/79/ES a NV SR č. 569/2001 Z. z. v platnom znení.

Dátum kontroly : 24.05.2022

Vedúca výrobného oddelenia:



Preveril:



BIOGEMA
výrobné družstvo, Košice
Garbiarska 2
040 01 KOŠICE ☎
tel.: 055/633 6753, 622 8131

O210605UE

