

OSVEDČENIE O SÚPRAVE

ELISA-anti-MILK-II-G-200

Katalógové číslo	REF	E-245/200
Šarža	LOT	05 05 22
Dátum výroby		24.05.2022
Dátum expirácie		24.02.2023

Parametre výstupnej kontroly

Štandardy		protilátky anti IgG
		Absorbancia pri 450 nm a 630 nm
0	U.ml ⁻¹ 	0,007
6,25	U.ml ⁻¹ 	0,260
12,5	U.ml ⁻¹ 	0,463
25	U.ml ⁻¹ 	0,722
50	U.ml ⁻¹ 	1,041
200	U.ml ⁻¹ 	1,630

Protilátky triedy IgG

Kontrolné vzorky	Deklarovaná hodnota U.ml ⁻¹	Absorbancia pri 450 a 630 nm	Nameraná hodnota U.ml ⁻¹
negatívna 	< 9	0,107	2,28
pozitívna 	29 ± 6	0,782	28,78

ZÁVER:

Na základe výstupnej kontroly šarža výrobku vyhovuje špecifikácii - technologický reglement č.7 , Smernici 98/79/ES a NV SR č. 569/2001 Z. z. v platnom znení.

Dátum kontroly : 24.05.2022

Vedúca výrobného oddelenia:



Preveril:



BIOGEMA
výrobné družstvo, Košice
Garbiarska 2
040 01 KOŠICE ☎
tel.: 055/633 6753, 622 8131

O210605YE

