

OSVEDČENIE O SÚPRAVE

ELISA-anti-CANDIDA-II-G

Katalógové číslo	REF	E-352
Šarža	LOT	07 07 22
Dátum výroby		21.07.2022
Dátum expirácie		21.07.2023

Parametre výstupnej kontroly

Štandardy	protilátky anti IgG	
	Absorbancia pri 450 nm a 630 nm	
0 U.ml ⁻¹		0,012
6,25 U.ml ⁻¹		0,806
12,5 U.ml ⁻¹		1,287
25 U.ml ⁻¹		1,600
50 U.ml ⁻¹		1,781
100 U.ml ⁻¹		1,928

Protilátky triedy IgG

Kontrolné vzorky	Deklarovaná hodnota U.ml ⁻¹	Absorbancia pri 450 a 630 nm	Nameraná hodnota U.ml ⁻¹
negatívna 	< 12	0,472	3,097
pozitívna 	40 ± 8	1,710	37,077

ZÁVER:

Na základe výstupnej kontroly šarža výrobku vyhovuje špecifikácii - technologický reglement č.26 , Smernici 98/79/ES a NV SR č. 569/2001 Z. z. v platnom znení.

Dátum kontroly : 21.07.2022

Vedúca výrobného oddelenia:


BIOGEMA
výrobné družstvo, Košice
Garbiarska 2
040 01 KOŠICE
tel.: 055/633 6753, 622 8131

Preveril:



R211222AG

