

OSVEDČENIE O SÚPRAVE

ELISA-anti-CANDIDA-II-M

Katalógové číslo	REF	E-353
Šarža	LOT	07 07 22
Dátum výroby		21.07.2022
Dátum expirácie		21.07.2023

Parametre výstupnej kontroly

Štandardy	protilátky anti IgM	
	Absorbancia pri 450 nm a 630 nm	
0 U.ml ⁻¹		0,009
6,25 U.ml ⁻¹		0,262
12,5 U.ml ⁻¹		0,493
25 U.ml ⁻¹		0,835
50 U.ml ⁻¹		1,310
100 U.ml ⁻¹		1,947

Protilátky triedy IgM

Kontrolné vzorky	Deklarovaná hodnota U.ml ⁻¹	Absorbancia pri 450 a 630 nm	Nameraná hodnota U.ml ⁻¹
negatívna 	< 12	0,264	6,299
pozitívna 	39 ± 7	1,097	37,632

ZÁVER:

Na základe výstupnej kontroly šarža výrobku vyhovuje špecifikácii - technologický reglement č.26 , Smernici 98/79/ES a NV SR č. 569/2001 Z. z. v platnom znení.

Dátum kontroly : 21.07.2022

Vedúca výrobného oddelenia:


BIOGEMA
výrobné družstvo, Košice
Garbiarska 2
040 01 KOŠICE
tel.: 055/633 6753, 622 8131

Preveril:



R211222RG

