

OSVEDČENIE O SÚPRAVE

ELISA-anti-MILK-II-G-200

Katalógové číslo	REF	E-245/200
Šarža	LOT	10 10 22
Dátum výroby		24.10.2022
Dátum expirácie		24.07.2023

Parametre výstupnej kontroly

Štandardy		protilátky anti IgG
		Absorbancia pri 450 nm a 630 nm
0	U.ml ⁻¹ 	0,004
6,25	U.ml ⁻¹ 	0,208
12,5	U.ml ⁻¹ 	0,377
25	U.ml ⁻¹ 	0,619
50	U.ml ⁻¹ 	0,942
200	U.ml ⁻¹ 	1,467

Protilátky triedy IgG

Kontrolné vzorky	Deklarovaná hodnota U.ml ⁻¹	Absorbancia pri 450 a 630 nm	Nameraná hodnota U.ml ⁻¹
negatívna 	< 9	0,059	1,52
pozitívna 	29 ± 6	0,711	30,83

ZÁVER:

Na základe výstupnej kontroly šarža výrobku vyhovuje špecifikácii - technologický reglement č.7 ,
Smernici 98/79/ES a NV SR č. 569/2001 Z. z. v platnom znení.

Dátum kontroly : 24.10.2022

Vedúca výrobného oddelenia:



Preveril:



BIOGEMA
výrobné družstvo, Košice
Garbiarska 2
040 01 KOŠICE ☎
tel.: 055/633 6753, 622 8131

O171018UK

