

OSVEDČENIE O SÚPRAVE

ELISA-anti-CANDIDA-II-A

Katalógové číslo	REF	E-351
Šarža	LOT	10 10 22
Dátum výroby		24.10.2022
Dátum expirácie		24.10.2023

Parametre výstupnej kontroly

Štandardy		protilátky anti IgA	
		Absorbancia pri 450 nm a 630 nm	
0	U.ml ⁻¹		0,006
6,25	U.ml ⁻¹		0,317
12,5	U.ml ⁻¹		0,603
25	U.ml ⁻¹		0,930
50	U.ml ⁻¹		1,321
100	U.ml ⁻¹		1,596

Protilátky triedy IgA

Kontrolné vzorky	Deklarovaná hodnota U.ml ⁻¹	Absorbancia pri 450 a 630 nm	Nameraná hodnota U.ml ⁻¹
negatívna 	< 8	0,272	5,361
pozitívna 	38 ± 7	1,158	37,420

ZÁVER:

Na základe výstupnej kontroly šarža výrobku vyhovuje špecifikácii - technologický reglement č.26 , Smernici 98/79/ES a NV SR č. 569/2001 Z. z. v platnom znení.

Dátum kontroly : 24.10.2022

Vedúca výrobného oddelenia:


BIOGEMA
výrobné družstvo, Košice
Garbiarska 2
040 01 KOŠICE
tel.: 055/633 6753, 622 8131

Preveril:



T171018CK

