

OSVEDČENIE O SÚPRAVE

ELISA-anti-CANDIDA-II-M

Katalógové číslo	REF	E-353
Šarža	LOT	10 10 22
Dátum výroby		24.10.2022
Dátum expirácie		24.10.2023

Parametre výstupnej kontroly

Štandardy	protilátky anti IgM	
	Absorbancia pri 450 nm a 630 nm	
0 U.ml ⁻¹		0,016
6,25 U.ml ⁻¹		0,269
12,5 U.ml ⁻¹		0,455
25 U.ml ⁻¹		0,805
50 U.ml ⁻¹		1,228
100 U.ml ⁻¹		1,726

Protilátky triedy IgM

Kontrolné vzorky	Deklarovaná hodnota U.ml ⁻¹	Absorbancia pri 450 a 630 nm	Nameraná hodnota U.ml ⁻¹
negatívna 	< 12	0,264	6,086
pozitívna 	39 ± 7	1,018	35,631

ZÁVER:

Na základe výstupnej kontroly šarža výrobku vyhovuje špecifikácii - technologický reglement č.26 , Smernici 98/79/ES a NV SR č. 569/2001 Z. z. v platnom znení.

Dátum kontroly : 24.10.2022

Vedúca výrobného oddelenia:


BIOGEMA
výrobné družstvo, Košice
Garbiarska 2
040 01 KOŠICE
tel.: 055/633 6753, 622 8131

Preveril:



R171018RK

