

OSVEDČENIE O SÚPRAVE

ELISA-anti-ASCA-II-A

Katalógové číslo	REF	E-346
Šarža	LOT	12 12 22
Dátum výroby		19.12.2022
Dátum expirácie		19.09.2023

Parametre výstupnej kontroly

Štandardy		protilátky anti IgA
		Absorbancia pri 450 nm a 630 nm
0	U.ml ⁻¹ 	0,011
6,25	U.ml ⁻¹ 	0,170
12,5	U.ml ⁻¹ 	0,343
25	U.ml ⁻¹ 	0,640
50	U.ml ⁻¹ 	1,011
100	U.ml ⁻¹ 	1,442

Protilátky triedy IgA

Kontrolné vzorky	Deklarovaná hodnota U.ml ⁻¹	Absorbancia pri 450 a 630 nm	Nameraná hodnota U.ml ⁻¹
negatívna 	< 10	0,110	4,08
pozitívna 	49 ± 9	1,001	49,07

ZÁVER:

Na základe výstupnej kontroly šarža výrobku vyhovuje špecifikácii - technologický reglement č.20 , Smernici 98/79/ES a NV SR č. 569/2001 Z. z. v platnom znení.

Dátum kontroly : 19.12.2022

Vedúca výrobného oddelenia:



Preveril:



BIOGEMA
výrobné družstvo, Košice
Garbiarska 2
040 01 KOŠICE
tel.: 055/633 6753, 622 8131