

OSVEDČENIE O SÚPRAVE

ELISA-anti-MILK-II-A

Katalógové číslo	REF	E-244
Šarža	LOT	02 02 23
Dátum výroby		22.02.2023
Dátum expirácie		22.11.2023

Parametre výstupnej kontroly

Štandardy		protilátky anti IgA
		Absorbancia pri 450 nm a 630 nm
0	U.ml ⁻¹ 	0,010
6,25	U.ml ⁻¹ 	0,222
12,5	U.ml ⁻¹ 	0,420
25	U.ml ⁻¹ 	0,701
50	U.ml ⁻¹ 	1,146
100	U.ml ⁻¹ 	1,589

Protilátky triedy IgA

Kontrolné vzorky	Deklarovaná hodnota U.ml ⁻¹	Absorbancia pri 450 a 630 nm	Nameraná hodnota U.ml ⁻¹
negatívna 	< 9	0,035	0,94
pozitívna 	29 ± 6	0,766	28,19

ZÁVER:

Na základe výstupnej kontroly šarža výrobku vyhovuje špecifikácii - technologický reglement č.5 , Smernici 98/79/ES a NV SR č. 569/2001 Z. z. v platnom znení.

Dátum kontroly : 22.02.2023

Vedúca výrobného oddelenia:



Preveril:



BIOGEMA
výrobné družstvo, Košice
Garbiarska 2
040 01 KOŠICE ☎
tel.: 055/633 6753, 622 8131

O200810UB

