

OSVEDČENIE O SÚPRAVE

ELISA-anti-ASCA-II-A

Katalógové číslo	REF	E-346
Šarža	LOT	02 02 23
Dátum výroby		22.02.2023
Dátum expirácie		22.11.2023

Parametre výstupnej kontroly

Štandardy		protilátky anti IgA
		Absorbancia pri 450 nm a 630 nm
0	U.ml ⁻¹ 	0,012
6,25	U.ml ⁻¹ 	0,178
12,5	U.ml ⁻¹ 	0,329
25	U.ml ⁻¹ 	0,604
50	U.ml ⁻¹ 	0,972
100	U.ml ⁻¹ 	1,471

Protilátky triedy IgA

Kontrolné vzorky	Deklarovaná hodnota U.ml ⁻¹	Absorbancia pri 450 a 630 nm	Nameraná hodnota U.ml ⁻¹
negatívna 	< 10	0,109	3,57
pozitívna 	48 ± 10	0,937	47,12

ZÁVER:

Na základe výstupnej kontroly šarža výrobku vyhovuje špecifikácii - technologický reglement č.20 , Smernici 98/79/ES a NV SR č. 569/2001 Z. z. v platnom znení.

Dátum kontroly : 22.02.2023

Vedúca výrobného oddelenia:



Preveril:



BIOGEMA
výrobné družstvo, Košice
Garbiarska 2
040 01 KOŠICE
tel.: 055/633 6753, 622 8131