

OSVEDČENIE O SÚPRAVE

ELISA-anti-CANDIDA-II-G

Katalógové číslo	REF	E-352
Šarža	LOT	02 02 23
Dátum výroby		22.02.2023
Dátum expirácie		22.02.2024

Parametre výstupnej kontroly

Štandardy	protilátky anti IgG	
	Absorbancia pri 450 nm a 630 nm	
0 U.ml ⁻¹		0,008
6,25 U.ml ⁻¹		0,742
12,5 U.ml ⁻¹		1,085
25 U.ml ⁻¹		1,282
50 U.ml ⁻¹		1,419
100 U.ml ⁻¹		1,524

Protilátky triedy IgG

Kontrolné vzorky	Deklarovaná hodnota U.ml ⁻¹	Absorbancia pri 450 a 630 nm	Nameraná hodnota U.ml ⁻¹
negatívna 	< 12	0,572	4,120
pozitívna 	39 ± 8	1,368	38,141

ZÁVER:

Na základe výstupnej kontroly šarža výrobku vyhovuje špecifikácii - technologický reglement č.26 , Smernici 98/79/ES a NV SR č. 569/2001 Z. z. v platnom znení.

Dátum kontroly : 22.02.2023

Vedúca výrobného oddelenia:


BIOGEMA
výrobné družstvo, Košice
Garbiarska 2
040 01 KOŠICE
tel.: 055/633 6753, 622 8131

Preveril:



R200810AB

