

OSVEDČENIE O SÚPRAVE

ELISA-anti-MILK-II-G

Katalógové číslo	REF	E-245
Šarža	LOT	04 04 23
Dátum výroby		24.04.2023
Dátum expirácie		24.01.2024

Parametre výstupnej kontroly

Štandardy		protilátky anti IgG
		Absorbancia pri 450 nm a 630 nm
0	U.ml ⁻¹ 	0,006
6,25	U.ml ⁻¹ 	0,359
12,5	U.ml ⁻¹ 	0,618
25	U.ml ⁻¹ 	0,963
50	U.ml ⁻¹ 	1,252
100	U.ml ⁻¹ 	1,630

Protilátky triedy IgG

Kontrolné vzorky	Deklarovaná hodnota U.ml ⁻¹	Absorbancia pri 450 a 630 nm	Nameraná hodnota U.ml ⁻¹
negatívna 	< 9	0,099	1,24
pozitívna 	30 ± 6	1,081	32,70

ZÁVER:

Na základe výstupnej kontroly šarža výrobku vyhovuje špecifikácii - technologický reglement č.7 ,
Smernici 98/79/ES a NV SR č. 569/2001 Z. z. v platnom znení.

Dátum kontroly : 24.04.2023

Vedúca výrobného oddelenia:



Preveril:



BIOGEMA
výrobné družstvo, Košice
Garbiarska 2
040 01 KOŠICE ☎
tel.: 055/633 6753, 622 8131

O210226YD

