

OSVEDČENIE O SÚPRAVE

ELISA-anti-MILK-II-A-200

Katalógové číslo	REF	E-244/200
Šarža	LOT	04 04 23
Dátum výroby		24.04.2023
Dátum expirácie		24.01.2024

Parametre výstupnej kontroly

Štandardy	protilátky anti IgA	
	Absorbancia pri 450 nm a 630 nm	
0 U.ml ⁻¹		0,006
6,25 U.ml ⁻¹		0,131
12,5 U.ml ⁻¹		0,236
25 U.ml ⁻¹		0,418
50 U.ml ⁻¹		0,737
200 U.ml ⁻¹		1,643

Protilátky triedy IgA

Kontrolné vzorky	Deklarovaná hodnota U.ml ⁻¹	Absorbancia pri 450 a 630 nm	Nameraná hodnota U.ml ⁻¹
negatívna 	< 9	0,013	0,42
pozitívna 	29 ± 6	0,538	33,76

ZÁVER:

Na základe výstupnej kontroly šarža výrobku vyhovuje špecifikácii - technologický reglement č.5 , Smernici 98/79/ES a NV SR č. 569/2001 Z. z. v platnom znení.

Dátum kontroly : 24.04.2023

Vedúca výrobného oddelenia:



Preveril:



BIOGEMA
výrobné družstvo, Košice
Garbiarska 2
040 01 KOŠICE ☎
tel.: 055/633 6753, 622 8131

O200810UB

