

OSVEDČENIE O SÚPRAVE

ELISA-anti-MILK-II-M

Katalógové číslo	REF	E-243
Šarža	LOT	05 05 23
Dátum výroby		23.05.2023
Dátum expirácie		23.02.2024

Parametre výstupnej kontroly

Štandardy	protilátky anti IgM	
	Absorbancia pri 450 nm a 630 nm	
0 U.ml ⁻¹		0,010
6,25 U.ml ⁻¹		0,297
12,5 U.ml ⁻¹		0,597
25 U.ml ⁻¹		1,064
50 U.ml ⁻¹		1,532
100 U.ml ⁻¹		1,972

Protilátky triedy IgM

Kontrolné vzorky	Deklarovaná hodnota U.ml ⁻¹	Absorbancia pri 450 a 630 nm	Nameraná hodnota U.ml ⁻¹
negatívna 	< 9	0,148	3,19
pozitívna 	33 ± 7	1,301	35,07

ZÁVER:

Na základe výstupnej kontroly šarža výrobku vyhovuje špecifikácii - technologický reglement č.22 , Smernici 98/79/ES a NV SR č. 569/2001 Z. z. v platnom znení.

Dátum kontroly : 23.05.2023

Vedúca výrobného oddelenia:



Preveril:



BIOGEMA
výrobné družstvo, Košice
Garbiarska 2
040 01 KOŠICE ☎
tel.: 055/633 6753, 622 8131