









OSVEDČENIE O SÚPRAVE



ELISA-anti-MILK-II-A-200

Katalógové číslo	REF	E-244/200
Šarža	LOT	05 05 23
Dátum výroby		23.03.2023
Dátum expirácie		23.02.2024

Parametre výstupnej kontroly

Štandardy	protilátky anti IgA	
	Absorbancia pri 450 nm a 630 nm	
0 U.ml ⁻¹		0,006
6,25 U.ml ⁻¹		0,148
12,5 U.ml ⁻¹		0,261
25 U.ml ⁻¹		0,518
50 U.ml ⁻¹		0,877
200 U.ml ⁻¹		1,720

Protilátky triedy IgA

Kontrolné vzorky	Deklarovaná hodnota U.ml ⁻¹	Absorbancia pri 450 a 630 nm	Nameraná hodnota U.ml ⁻¹
negatívna 	< 9	0,020	0,45
pozitívna 	29 ± 6	0,580	28,44

ZÁVER:

Na základe výstupnej kontroly šarža výrobku vyhovuje špecifikácii - technologický reglement č.5 , Smernici 98/79/ES a NV SR č. 569/2001 Z. z. v platnom znení.

Dátum kontroly : 23.05.2023

Vedúca výrobného oddelenia:



Preveril:



BIOGEMA
výrobné družstvo, Košice
Garbiarska 2
040 01 KOŠICE ☐
tel.: 055/633 6753, 622 8131

O210606UE

