

OSVEDČENIE O SÚPRAVE

ELISA-anti-ASCA-II-A

Katalógové číslo	REF	E-346
Šarža	LOT	05 05 23
Dátum výroby		23.05.2023
Dátum expirácie		23.02.2024

Parametre výstupnej kontroly

Štandardy		protilátky anti IgA
		Absorbancia pri 450 nm a 630 nm
0	U.ml ⁻¹ 	0,012
6,25	U.ml ⁻¹ 	0,204
12,5	U.ml ⁻¹ 	0,405
25	U.ml ⁻¹ 	0,681
50	U.ml ⁻¹ 	1,157
100	U.ml ⁻¹ 	1,594

Protilátky triedy IgA

Kontrolné vzorky	Deklarovaná hodnota U.ml ⁻¹	Absorbancia pri 450 a 630 nm	Nameraná hodnota U.ml ⁻¹
negatívna 	< 10	0,132	4,14
pozitívna 	48 ± 10	1,129	48,13

ZÁVER:

Na základe výstupnej kontroly šarža výrobku vyhovuje špecifikácii - technologický reglement č.20 , Smernici 98/79/ES a NV SR č. 569/2001 Z. z. v platnom znení.

Dátum kontroly : 23.05.2023

Vedúca výrobného oddelenia:



Preveril:



BIOGEMA
výrobné družstvo, Košice
Garbiarska 2
040 01 KOŠICE
tel.: 055/633 6753, 622 8131