

OSVEDČENIE O SÚPRAVE

ELISA-anti-CANDIDA-II-A

Katalógové číslo	REF	E-351
Šarža	LOT	06 06 23
Dátum výroby		23.06.2023
Dátum expirácie		23.06.2024

Parametre výstupnej kontroly

Štandardy		protilátky anti IgA	
		Absorbancia pri 450 nm a 630 nm	
0	U.ml ⁻¹		0,009
6,25	U.ml ⁻¹		0,243
12,5	U.ml ⁻¹		0,434
25	U.ml ⁻¹		0,692
50	U.ml ⁻¹		1,078
100	U.ml ⁻¹		1,422

Protilátky triedy IgA

Kontrolné vzorky	Deklarovaná hodnota U.ml ⁻¹	Absorbancia pri 450 a 630 nm	Nameraná hodnota U.ml ⁻¹
negatívna 	< 8	0,201	5,032
pozitívna 	42 ± 8	0,962	41,037

ZÁVER:

Na základe výstupnej kontroly šarža výrobku vyhovuje špecifikácii - technologický reglement č.26 , Smernici 98/79/ES a NV SR č. 569/2001 Z. z. v platnom znení.

Dátum kontroly : 23.06.2023

Vedúca výrobného oddelenia:


BIOGEMA
výrobné družstvo, Košice
Garbiarska 2
040 01 KOŠICE
tel.: 055/633 6753, 622 8131

Preveril:



T210914CF

