

OSVEDČENIE O SÚPRAVE

ELISA-anti-MILK-II-G-200

Katalógové číslo	REF	E-245/200
Šarža	LOT	07 07 23
Dátum výroby		21.07.2023
Dátum expirácie		21.04.2024

Parametre výstupnej kontroly

Štandardy		protilátky anti IgG
		Absorbancia pri 450 nm a 630 nm
0	U.ml ⁻¹ 	0,005
6,25	U.ml ⁻¹ 	0,232
12,5	U.ml ⁻¹ 	0,429
25	U.ml ⁻¹ 	0,717
50	U.ml ⁻¹ 	1,056
200	U.ml ⁻¹ 	1,621

Protilátky triedy IgG

Kontrolné vzorky	Deklarovaná hodnota U.ml ⁻¹	Absorbancia pri 450 a 630 nm	Nameraná hodnota U.ml ⁻¹
negatívna 	< 9	0,077	1,88
pozitívna 	29 ± 6	0,834	31,83

ZÁVER:

Na základe výstupnej kontroly šarža výrobku vyhovuje špecifikácii - technologický reglement č.7 , Smernici 98/79/ES a NV SR č. 569/2001 Z. z. v platnom znení.

Dátum kontroly : 21.07.2023

Vedúca výrobného oddelenia:



Preveril:



BIOGEMA
výrobné družstvo, Košice
Garbiarska 2
040 01 KOŠICE ☎
tel.: 055/633 6753, 622 8131

O211223YG

