









OSVEDČENIE O SÚPRAVE



ELISA-anti-MILK-II-M

Katalógové číslo	REF	E-243
Šarža	LOT	10 10 23
Dátum výroby		23.10.2023
Dátum expirácie		23.07.2024

Parametre výstupnej kontroly

Štandardy	protilátky anti IgM	
	Absorbancia pri 450 nm a 630 nm	
0 U.ml ⁻¹ 	0,013	
6,25 U.ml ⁻¹ 	0,421	
12,5 U.ml ⁻¹ 	0,759	
25 U.ml ⁻¹ 	1,154	
50 U.ml ⁻¹ 	1,630	
100 U.ml ⁻¹ 	2,013	

Protilátky triedy IgM

Kontrolné vzorky	Deklarovaná hodnota U.ml ⁻¹	Absorbancia pri 450 a 630 nm	Nameraná hodnota U.ml ⁻¹
negatívna 	< 9	0,278	3,96
pozitívna 	20 - 50	1,323	32,21

ZÁVER:

Na základe výstupnej kontroly šarža výrobku vyhovuje špecifikácii - technologický reglement č.22 , Smernici 98/79/ES a NV SR č. 569/2001 Z. z. v platnom znení.

Dátum kontroly : 23.10.2023

Vedúca výrobného oddelenia:



Preveril:



BIOGEMA
výrobné družstvo, Košice
Garbiarska 2
040 01 KOŠICE ☎
tel.: 055/633 6753, 622 8131