

OSVEDČENIE O SÚPRAVE

ELISA-anti-MILK-II-G-200

Katalógové číslo	REF	E-245/200
Šarža	LOT	09 09 23
Dátum výroby		22.09.2023
Dátum expirácie		22.06.2024

Parametre výstupnej kontroly

Štandardy		protilátky anti IgG
		Absorbancia pri 450 nm a 630 nm
0	U.ml ⁻¹ 	0,005
6,25	U.ml ⁻¹ 	0,340
12,5	U.ml ⁻¹ 	0,578
25	U.ml ⁻¹ 	0,823
50	U.ml ⁻¹ 	1,116
200	U.ml ⁻¹ 	1,733

Protilátky triedy IgG

Kontrolné vzorky	Deklarovaná hodnota U.ml ⁻¹	Absorbancia pri 450 a 630 nm	Nameraná hodnota U.ml ⁻¹
negatívna 	< 9	0,135	2,07
pozitívna 	20 - 50	0,902	30,55

ZÁVER:

Na základe výstupnej kontroly šarža výrobku vyhovuje špecifikácii - technologický reglement č.7 , Smernici 98/79/ES a NV SR č. 569/2001 Z. z. v platnom znení.

Dátum kontroly : 22.09.2023

Vedúca výrobného oddelenia:



Preveril:



BIOGEMA
výrobné družstvo, Košice
Garbiarska 2
040 01 KOŠICE ☎
tel.: 055/633 6753, 622 8131

O220711YI

