

OSVEDČENIE O SÚPRAVE

ELISA-anti-MILK-II-G-200

Katalógové číslo	REF	E-245/200
Šarža	LOT	11 11 23
Dátum výroby		24.11.2023
Dátum expirácie		24.08.2024

Parametre výstupnej kontroly

Štandardy	protilátky anti IgG	
	Absorbancia pri 450 nm a 630 nm	
0 U.ml ⁻¹		0,005
6,25 U.ml ⁻¹		0,248
12,5 U.ml ⁻¹		0,464
25 U.ml ⁻¹		0,714
50 U.ml ⁻¹		1,017
200 U.ml ⁻¹		1,625

Protilátky triedy IgG

Kontrolné vzorky	Deklarovaná hodnota U.ml ⁻¹	Absorbancia pri 450 a 630 nm	Nameraná hodnota U.ml ⁻¹
negatívna 	< 9	0,122	3,03
pozitívna 	20 - 50	0,793	30,42

ZÁVER:

Na základe výstupnej kontroly šarža výrobku vyhovuje špecifikácii - technologický reglement č.7 , Smernici 98/79/ES a NV SR č. 569/2001 Z. z. v platnom znení.

Dátum kontroly : 24.11.2023

Vedúca výrobného oddelenia:



Preveril:



BIOGEMA
výrobné družstvo, Košice
Garbiarska 2
040 01 KOŠICE ☎
tel.: 055/633 6753, 622 8131

O230127YL

