

OSVEDČENIE O SÚPRAVE

ELISA-anti-CANDIDA-II-A

Katalógové číslo	REF	E-351
Šarža	LOT	12 12 23
Dátum výroby		19.12.2023
Dátum expirácie		19.12.2024

Parametre výstupnej kontroly

Štandardy	protilátky anti IgA	
	Absorbancia pri 450 nm a 630 nm	
0 U.ml ⁻¹		0,005
6,25 U.ml ⁻¹		0,209
12,5 U.ml ⁻¹		0,383
25 U.ml ⁻¹		0,676
50 U.ml ⁻¹		1,008
100 U.ml ⁻¹		1,392

Protilátky triedy IgA

Kontrolné vzorky	Deklarovaná hodnota U.ml ⁻¹	Absorbancia pri 450 a 630 nm	Nameraná hodnota U.ml ⁻¹
negatívna 	< 8	0,218	6,577
pozitívna 	20 - 50	0,886	38,866

ZÁVER:

Na základe výstupnej kontroly šarža výrobku vyhovuje špecifikácii - technologický reglement č.26 , Smernici 98/79/ES a NV SR č. 569/2001 Z. z. v platnom znení.

Dátum kontroly : 19.12.2023

Vedúci výrobného oddelenia:


BIOGEMA
výrobné družstvo, Košice
Garbiarska 2
040 01 KOŠICE ☎
tel.: 055/633 6753, 622 8131

Preveril:



T230507CM

