

OSVEDČENIE O SÚPRAVE

ELISA-anti-MILK-II-M

Katalógové číslo	REF	E-243
Šarža	LOT	01 01 26
Dátum výroby		23.01.2026
Dátum expirácie		23.10.2026

Parametre výstupnej kontroly

Štandardy	protilátky anti IgM	
	Absorbancia pri 450 nm a 630 nm	
0 U.ml ⁻¹		0,007
6,25 U.ml ⁻¹		0,215
12,5 U.ml ⁻¹		0,405
25 U.ml ⁻¹		0,704
50 U.ml ⁻¹		1,092
100 U.ml ⁻¹		1,567

Protilátky triedy IgM

Kontrolné vzorky	Deklarovaná hodnota U.ml ⁻¹	Absorbancia pri 450 a 630 nm	Nameraná hodnota U.ml ⁻¹
negatívna 	< 9	0,123	3,43
pozitívna 	20 - 50	0,825	31,47

ZÁVER:

Na základe výstupnej kontroly šarža výrobku vyhovuje špecifikácii - technologický reglement č.22 , Smernici 98/79/ES a Zákonu SR č. 56/2018 Z. z. v platnom znení.

Dátum kontroly : 23.01.2026

Vedúci výrobného oddelenia:

Preveril:


BIOGEMA
výrobné družstvo, Košice
Garbiarska 2
040 01 K O Š I C E
tel.: 055/633 6753, 622 8131

